

MacroMedics ® ThoraxTilt





MacroMedics BV
Oostbaan 670
2841ML Moordrecht
The Netherlands
+31(0)182 389 777
info@macromedics.com
www.macromedics.com



[This page is intentionally left blank]



Table of Contents

1.	Warnings	4
	MRI Safety Information	
	Intended Use	
	Installation Instructions	
5.	Cleaning, Storage and Disposal	6
6.	Warranty	6
7.	Notices	6
۵	Glossary of Symbols	7



1. Warnings



- Inspect the device thoroughly before use. Do not use the device if damaged or if missing parts.
- Check individual components on MR-safety when used together in a patient positioning system.
- Place and fix the device properly before positioning the patient.
- Do not exceed an equally distributed load of 75 kilograms.
- Ensure that the baseplate is correctly secured to the couch prior to use.
- In the USA, federal law restricts this device to sale by or on the order of a licensed healthcare practitioner.

2. MRI Safety Information

2.1 MR SAFE

The following devices are MR safe:

115420 ThoraxTilt Set MB




3. Intended Use

Positioning of the patient during radiotherapy and radiodiagnostics, including MR where indicated.



4. Installation Instructions

The ThoraxTilt consists of three stackable layers and is compatible with the MultiBoard. Place the Lowered part, at the bottom of the ThoraxTilt, into the cushion insert of the MultiBoard so that the ThoraxTilt is secured to the MultiBoard.





5. Cleaning, Storage and Disposal

- Clean the device before and after every use with 70% Isopropanol or soapy water, applied with a soft cloth. Rinse with water and dry properly. Follow the cleaning precautions detailed on the product provided by the manufacturer of the cleaning agent.
- If unsure about the cleaning fluid, do not use it. Never use Ethanol (Ethyl alcohol), aerosol sprays, corrosive cleaning agents, solvents or abrasive detergents.
- Make sure that there is no residue of the cleaning agent left on the device prior to use.
- The devices are not suitable for sterilization or thermal disinfection.
- Store the device in a clean, ventilated environment.
- Dispose of the device in accordance with the local laws and regulations applicable to your country.

6. Warranty

The MacroMedics products are warranted to be free from defects in manufacturing, material and construction. The warranty is for 1 year starting from the date of delivery.

Product warranties remain valid provided the product was installed and/or used according to the instructions. Defects, malfunctions or failures caused by nature, misuse, abuse and unauthorised alteration or repair are not warranted.

This warranty is limited to the repair and/or replacement at the sole discretion of MacroMedics.

7. Notices

Any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user or patient is established.



8. Glossary of Symbols

Symbol	Symbol Name	Description of Symbol
CE	European Conformity	EC declaration of conformity by manufacturer.
***	Manufacturer	Indicates the medical device manufacturer.
REF	Catalogue Number	Indicates the manufacturer's catalogue number so that the medical device can be identified.
LOT GS1 Code: (10)	Batch Code	Indicates the manufacturer's batch code so that the batch or lot can be identified.
[]i	Consult Instruction for Use	Indicates the need for the user to consult the instructions for use.
	Caution	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.
Rx ONLY	Prescription Only	Devices for sale to physicians or trained healthcare providers only.
Ť	Keep Dry	Indicates a medical device that needs to be protected from moisture.
	Keep Away from Sunlight	Indicates a medical device that needs protection from light and heat sources.
MR	MR Safe	An item that poses no known hazards resulting from exposure to any MR environment. MR Safe items are composed of materials that are electrically nonconductive, nonmetallic, and nonmagnetic.



MacroMedics BV
Oostbaan 670

2841ML Moordrecht The Netherlands



+31(0)182 389 777 info@macromedics.com

www.macromedics.com

(Visit www.macromedics.com for a digital copy of this manual)

